



MEMORANDUM

TO: Mr. Addison Rice
Anderson, Mulholland and Associates

DATE: July 8, 2015

FROM: R. Infante 

FILE: 1412152AR1

checked by 
7/14/15

RE: Data Validation
Air samples
SDG: 1412152AR1

SUMMARY

Full validation was performed on the data for several gas samples analyzed for selected volatile organic compounds by method Compendium Method TO-15: Determination Of Volatile Organic Compounds (VOCs) In Air Collected In Specially-Prepared Canisters And Analyzed By Gas Chromatography/Mass Spectrometry (GC/MS), January, 1999. The samples were collected at the Bristol Myer Squib-Building 5 VI facility, Humacao, PR site on December 09, 2014 and submitted to Eurofins Air Toxics, Inc. of Folsom, California that analyzed and reported the results under delivery groups (SDG) 1412152AR1.

The sample results were assessed according to USEPA data validation guidance documents in the following order of precedence: Compendium Method TO-15. Determination Of Volatile Organic Compounds (VOCs) In Air Collected In Specially-Prepared Canisters And Analyzed By Gas Chromatography/Mass Spectrometry (GC/MS), January, 1999; Validating Air Samples. Volatile Organic Analysis of Ambient Air in Canisters by Method TO-15, (SOP # HW-31. Revision #4. October, 2006. The QC criteria and data validation actions listed on the data review worksheets are from the primary guidance document, unless otherwise noted. In general the data is valid as reported and may be used for decision making purposes.

The data results are acceptable for use. The following results were qualified as estimated (J): 1,2,4-trichlorobenzene in all samples due to the % RSD for the calibration factor outside method performance criteria. Results for Ethanol, 1,2-dichlorobenzene, 1,3-dichlorobenzene, 1,4-dichlorobenzene, alpha-chlorotoluene and Hexachlorobutadiene are qualified as estimated (J) in all sample due to % D in continuing calibration outside method performance criteria. Detected results for Toluene, 2-Propanol, and Acetone in samples B5IA-6 (2014), B5IA-7 (2014), B5IA-2 (2014) and B5IA-1 (2014) exceed the instrument calibration range and are considered estimated values. Detected results for Toluene, 2-Propanol, 4-Methyl-2-pentanone and Acetone in sample B5IA-4 (2014) exceed the instrument calibration range and are considered estimated values. Result qualified as an estimated value (J) by the validator and qualified (E) by the laboratory.

The laboratory issued the following statement:

"The work order was re-issued on July 1, 2015 for the following reasons:

1. To report additional compounds per client's request. While the initial report met the laboratory data quality requirements for the originally requested compounds, the additional compounds were not evaluated for quality compliance at the time of sample analysis. As a result, the re-issued report contains qualified data for several of the added compounds.

2. All Quality Control Limit exceedances and affected sample results are noted by flags. Each flag is defined at the bottom of this Case Narrative and on each Sample Result Summary page.

3. To report estimated values for target compound hits that are below the Reporting Limit but greater than the Method Detection Limit. Concentrations that are below the level at which the canister was certified may be false positives."

SAMPLES

The samples included in the review are listed below

Client Sample ID	Lab. Sample ID	Collected Date	Matrix	Analysis
B5IA-8 (2014)	1412152AR1-01A	12/09/2014	Air	VOCs
B5IA-10 (2014)	1412152AR1-02A	12/09/2014	Air	VOCs
B5IA-6 (2014)	1412152AR1-03A	12/09/2014	Air	VOCs
B5IA-7 (2014)	1412152AR1-04A	12/09/2014	Air	VOCs
B5IA-2 (2014)	1412152AR1-05A	12/09/2014	Air	VOCs
B5IA-1 (2014)	1412152AR1-06A	12/09/2014	Air	VOCs
B5IA-AA (2014)	1412152AR1-07A	12/09/2014	Air	VOCs
B5IA-4 (2014)	1412152AR1-08A	12/09/2014	Air	VOCs

REVIEW ELEMENTS

Sample data were reviewed for the following parameters, where applicable to the method

- o Agreement of analysis conducted with chain of custody (COC) form
- o Holding time and sample preservation
- o Gas chromatography/mass spectrometry (GC/MS) tunes
- o Initial and continuing calibrations
- o Method blanks/trip blanks/field blank
- o Canister cleaning certification criteria
- o Surrogate spike recovery
- o Internal standard performance and retention times
- o Field duplicate results
- o Laboratory control sample/laboratory control sample duplicate (LCS/LCSD) results
- o Quantitation limits and sample results

DISCUSSION

Agreement of Analysis Conducted with COC Request

Sample reports corresponded to the analytical request designated on the chain-of-custody form.

Holding Times and Sample Preservation

Sample preservation was acceptable.

Samples analyzed within method recommended holding time.

GC/MS Tunes

The frequency and abundance of bromofluorobenzene (BFB) tunes were within the QC acceptance criteria. All samples were analyzed within the tuning criteria associated with the method.

Initial and Continuing Calibrations

VOCs (Method TO-15)

The percent relative standard deviations (%RSDs) and response factors (RFs) of all target analytes were within the QC acceptance criteria in the initial calibration except for the compounds listed below. Correlation coefficients (r^2) of target analytes were within the QC acceptance criteria. Ongoing accuracy of the instrument was determined by the analysis of a continuing calibration standard. The following compounds did not meet the method calibration/continued calibration criteria:

DATE	LAB ID#	FILE	CRITERIA OUT RFs, %RSD, %D, r	COMPOUND	SAMPLES AFFECTED
Initial and continuing calibration met the method performance criteria except for the followings:					
10/06/14	V14L1006D		34 % RSD	1,2,4-trichlorobenzene	All samples.
12/16/14	V121602.d		38 % D	Ethanol	
			38 % D	1,2-dichlorobenzene	
			32 % D	1,3-dichlorobenzene	
			37 % D	1,4-dichlorobenzene	
			33 % D	alpha-chlorotoluene	
			39 % D	Hexachlorobutadiene	

Results were qualified as estimated (J) in all samples.

Method Blank/Trip Blank/Field Blank

Target analytes were not detected in laboratory method blanks except for the followings:

DATE ANALYZED	LAB ID	LEVEL/MATRIX	COMPOUND	CONCENTRATION/UNIT
12/16/14	1412152AR1-09A	Air/low	Hexane	0.10 ppbv
			1,4-dichlorobenzene	0.038 ppbv
			alpha-chlorotoluene	0.10 ppbv
			1,2-dichlorobenzene	0.046 ppbv
			1,2,4-trichlorobenzene	0.12 ppbv
			Hexachlorobutadiene	0.043ppbv

No action taken, 5x concentration in blank < the concentration found in samples.

Summa canister met cleaning certification criteria.

No trip/field blank analyzed with this data package.

Surrogate Spike Recovery

The surrogate recoveries were within the laboratory QC acceptance limits in all samples analyzed.

Internal Standard Performance

VOCs and Methanol (TO-15)

Samples were spiked with the method specified internal standard. Internal standard are performance and retention times met the QC acceptance criteria in all sample analyses and calibration standards.

Laboratory/Field Duplicate Results

Laboratory duplicates (LCS/LCSD) were analyzed as part of this data set. Target analytes meet the RPD performance criteria of + 25 % for analytes 5 x SQL.

LCS/LCSD Results

VOCs

LCS/LCSD (blank spike) associated with this data package were analyzed by the laboratory. Recoveries and RPD within laboratory control limits except for the following:

LCS ID	COMPOUND	% R	QC LIMIT
1412152AR1-11A/-11AA	1,2,4-trimethylbenzene	143%/142%	70 - 130
	alpha-chlorotoluene	167%/159%	70 - 130
	Ethanol	149%/152%	70 - 130

No action taken, professional judgment.

Quantitation Limits and Sample Results

Dilution was performed on sample B5IA-1 (2014) due to the presence of high level target species. All other samples diluted by a factor of less than 2.

Detected results for Toluene, 2-Propanol, and Acetone in samples B5IA-6 (2014), B5IA-7 (2014), B5IA-2 (2014) and B5IA-1 (2014) exceed the instrument calibration range and are considered estimated values.


Detected results for Toluene, 2-Propanol, 4-Methyl-2-pentanone and Acetone in sample B5IA-4 (2014) exceed the instrument calibration range and are considered estimated values.

Calculations were spot checked.

Certification

The following samples 1412152AR1-01A; 1412152AR1-02A; 1412152AR1-03A; 1412152AR1-04A; 1412152AR1-05A, 1412152AR1-06A, 1412152AR1-07A and 1412152AR1-08A were analyzed following standard procedures accepted by regulatory agencies. The quality control requirements met the methods criteria except in the occasions described in this document. The results are valid. Some of the results were qualified.


Rafael Infante
Chemist License 1888





Air Toxics


Client Sample ID: B5IA- 8 (2014)

Lab ID#: 1412152AR1-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121616r1	Date of Collection:	12/9/14 11:26:00 AM
Dil. Factor:	1.73	Date of Analysis:	12/16/14 07:01 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.86	0.46 J	4.3	2.3 J
Freon 114	0.17	Not Detected	1.2	Not Detected
Chloromethane	0.86	0.85 J	1.8	1.8 J
Vinyl Chloride	0.17	Not Detected	0.44	Not Detected
1,3-Butadiene	0.17	0.056 J	0.38	0.12 J
Bromomethane	0.86	Not Detected	3.4	Not Detected
Chloroethane	0.86	Not Detected	2.3	Not Detected
Freon 11	0.17	0.22	0.97	1.2
Ethanol	0.86	68 J0	1.6	130 J0
Freon 113	0.17	0.053 J	1.3	0.41 J
1,1-Dichloroethene	0.17	Not Detected	0.68	Not Detected
Acetone	0.86	45	2.0	100
2-Propanol	0.86	30	2.1	73
Carbon Disulfide	0.86	Not Detected	2.7	Not Detected
3-Chloropropene	0.86	Not Detected	2.7	Not Detected
Methylene Chloride	0.35	10	1.2	36
Methyl tert-butyl ether	0.17	0.25	0.62	0.91
trans-1,2-Dichloroethene	0.17	Not Detected	0.68	Not Detected
Hexane	0.17	0.16 J	0.61	0.58 J
1,1-Dichloroethane	0.17	Not Detected	0.70	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.86	0.61 J	2.6	1.8 J
cis-1,2-Dichloroethene	0.17	Not Detected	0.68	Not Detected
Tetrahydrofuran	0.86	0.70 J	2.6	2.1 J
Chloroform	0.17	0.049 J	0.84	0.24 J
1,1,1-Trichloroethane	0.17	Not Detected	0.94	Not Detected
Cyclohexane	0.17	Not Detected	0.60	Not Detected
Carbon Tetrachloride	0.17	0.080 J	1.1	0.50 J
2,2,4-Trimethylpentane	0.86	0.55 J	4.0	2.6 J
Benzene	0.17	0.13 J	0.55	0.42 J
1,2-Dichloroethane	0.17	Not Detected	0.70	Not Detected
Heptane	0.17	0.43	0.71	1.8
Trichloroethene	0.17	Not Detected	0.93	Not Detected
1,2-Dichloropropane	0.17	Not Detected	0.80	Not Detected
1,4-Dioxane	0.17	Not Detected	0.62	Not Detected
Bromodichloromethane	0.17	Not Detected	1.2	Not Detected
cis-1,3-Dichloropropene	0.17	Not Detected	0.78	Not Detected
4-Methyl-2-pentanone	0.17	10	0.71	43
Toluene	0.17	21	0.65	80
trans-1,3-Dichloropropene	0.17	Not Detected	0.78	Not Detected
1,1,2-Trichloroethane	0.17	Not Detected	0.94	Not Detected
Tetrachloroethene	0.17	Not Detected	1.2	Not Detected
2-Hexanone	0.86	Not Detected	3.5	Not Detected





Air Toxics

Client Sample ID: B5IA- 8 (2014)

Lab ID#: 1412152AR1-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121616r1	Date of Collection:	12/9/14 11:26:00 AM
Dil. Factor:	1.73	Date of Analysis:	12/16/14 07:01 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.17	Not Detected	1.5	Not Detected
1,2-Dibromoethane (EDB)	0.17	Not Detected	1.3	Not Detected
Chlorobenzene	0.17	Not Detected	0.80	Not Detected
Ethyl Benzene	0.17	2.7	0.75	12
m,p-Xylene	0.17	10	0.75	44
o-Xylene	0.17	1.0	0.75	4.5
Styrene	0.17	0.15 J	0.74	0.65 J
Bromoform	0.17	Not Detected	1.8	Not Detected
Cumene	0.17	Not Detected	0.85	Not Detected
1,1,2,2-Tetrachloroethane	0.17	Not Detected	1.2	Not Detected
Propylbenzene	0.17	Not Detected	0.85	Not Detected
4-Ethyltoluene	0.17	0.078 J	0.85	0.38 J
1,3,5-Trimethylbenzene	0.17	Not Detected	0.85	Not Detected
1,2,4-Trimethylbenzene	0.17	0.17 J	0.85	0.84 J
1,3-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,4-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
alpha-Chlorotoluene	0.86	Not Detected UJ	4.5	Not Detected UJ
1,2-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,2,4-Trichlorobenzene	0.86	Not Detected	6.4	Not Detected
Hexachlorobutadiene	0.86	Not Detected UJ	9.2	Not Detected UJ
Naphthalene	0.86	Not Detected	4.5	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	111	70-130
Toluene-d8	106	70-130
4-Bromofluorobenzene	98	70-130





Air Toxics


Client Sample ID: BSIA-10 (2014)

Lab ID#: 1412152AR1-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121617r1	Date of Collection:	12/9/14 11:48:00 AM
Dil. Factor:	1.80	Date of Analysis:	12/16/14 07:43 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.90	0.48 J	4.4	2.4 J
Freon 114	0.18	Not Detected	1.2	Not Detected
Chloromethane	0.90	0.78 J	1.8	1.6 J
Vinyl Chloride	0.18	Not Detected	0.46	Not Detected
1,3-Butadiene	0.18	Not Detected	0.40	Not Detected
Bromomethane	0.90	Not Detected	3.5	Not Detected
Chloroethane	0.90	Not Detected	2.4	Not Detected
Freon 11	0.18	0.21	1.0	1.2
Ethanol	0.90	53 J0	1.7	100 J0
Freon 113	0.18	0.058 J	1.4	0.44 J
1,1-Dichloroethene	0.18	Not Detected	0.71	Not Detected
Acetone	0.90	36	2.1	86
2-Propanol	0.90	25	2.2	61
Carbon Disulfide	0.90	1.8	2.8	5.4
3-Chloropropene	0.90	Not Detected	2.8	Not Detected
Methylene Chloride	0.36	8.0	1.2	28
Methyl tert-butyl ether	0.18	0.22	0.65	0.80
trans-1,2-Dichloroethene	0.18	Not Detected	0.71	Not Detected
Hexane	0.18	0.16 J	0.63	0.56 J
1,1-Dichloroethane	0.18	Not Detected	0.73	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.90	0.59 J	2.6	1.7 J
cis-1,2-Dichloroethene	0.18	Not Detected	0.71	Not Detected
Tetrahydrofuran	0.90	0.45 J	2.6	1.3 J
Chloroform	0.18	0.045 J	0.88	0.22 J
1,1,1-Trichloroethane	0.18	Not Detected	0.98	Not Detected
Cyclohexane	0.18	Not Detected	0.62	Not Detected
Carbon Tetrachloride	0.18	Not Detected	1.1	Not Detected
2,2,4-Trimethylpentane	0.90	0.53 J	4.2	2.5 J
Benzene	0.18	0.16 J	0.58	0.51 J
1,2-Dichloroethane	0.18	Not Detected	0.73	Not Detected
Heptane	0.18	0.38	0.74	1.5
Trichloroethene	0.18	Not Detected	0.97	Not Detected
1,2-Dichloropropane	0.18	Not Detected	0.83	Not Detected
1,4-Dioxane	0.18	0.28	0.65	1.0
Bromodichloromethane	0.18	Not Detected	1.2	Not Detected
cis-1,3-Dichloropropene	0.18	Not Detected	0.82	Not Detected
4-Methyl-2-pentanone	0.18	8.2	0.74	34
Toluene	0.18	17	0.68	64
trans-1,3-Dichloropropene	0.18	Not Detected	0.82	Not Detected
1,1,2-Trichloroethane	0.18	Not Detected	0.98	Not Detected
Tetrachloroethene	0.18	Not Detected	1.2	Not Detected
2-Hexanone	0.90	Not Detected	3.7	Not Detected





Air Toxics

Client Sample ID: B5IA-10 (2014)

Lab ID#: 1412152AR1-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121617r1	Date of Collection:	12/9/14 11:48:00 AM
Dil. Factor:	1.80	Date of Analysis:	12/16/14 07:43 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.18	Not Detected	1.5	Not Detected
1,2-Dibromoethane (EDB)	0.18	Not Detected	1.4	Not Detected
Chlorobenzene	0.18	Not Detected	0.83	Not Detected
Ethyl Benzene	0.18	2.2	0.78	9.5
m,p-Xylene	0.18	7.7	0.78	34
o-Xylene	0.18	0.92	0.78	4.0
Styrene	0.18	0.13 J	0.77	0.57 J
Bromoform	0.18	Not Detected	1.9	Not Detected
Cumene	0.18	Not Detected	0.88	Not Detected
1,1,2,2-Tetrachloroethane	0.18	Not Detected	1.2	Not Detected
Propylbenzene	0.18	Not Detected	0.88	Not Detected
4-Ethyltoluene	0.18	0.098 J	0.88	0.48 J
1,3,5-Trimethylbenzene	0.18	Not Detected	0.88	Not Detected
1,2,4-Trimethylbenzene	0.18	0.14 J	0.88	0.70 J
1,3-Dichlorobenzene	0.18	Not Detected UJ J	1.1	Not Detected UJ
1,4-Dichlorobenzene	0.18	0.027 J J	1.1	0.16 J
alpha-Chlorotoluene	0.90	0.11 J J	4.6	0.56 J
1,2-Dichlorobenzene	0.18	Not Detected UJ J	1.1	Not Detected UJ
1,2,4-Trichlorobenzene	0.90	Not Detected J	6.7	Not Detected
Hexachlorobutadiene	0.90	Not Detected UJ J	9.6	Not Detected UJ
Naphthalene	0.90	Not Detected	4.7	Not Detected


J = Estimated value.

J0 = Estimated value due to bias in the CCV.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	113	70-130
Toluene-d8	105	70-130
4-Bromofluorobenzene	96	70-130





Air Toxics


Client Sample ID: B5IA-6 (2014)

Lab ID#: 1412152AR1-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121618r1	Date of Collection:	12/9/14 1:29:00 PM
Dil. Factor:	1.58	Date of Analysis:	12/16/14 08:41 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.79	0.46 J	3.9	2.3 J
Freon 114	0.16	Not Detected	1.1	Not Detected
Chloromethane	0.79	0.89	1.6	1.8
Vinyl Chloride	0.16	Not Detected	0.40	Not Detected
1,3-Butadiene	0.16	Not Detected	0.35	Not Detected
Bromomethane	0.79	Not Detected	3.1	Not Detected
Chloroethane	0.79	Not Detected	2.1	Not Detected
Freon 11	0.16	0.18	0.89	1.0
Ethanol	0.79	280 JOE J	1.5	520 JOE
Freon 113	0.16	0.050 J	1.2	0.38 J
1,1-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Acetone	0.79	210 E J	1.9	510 E
2-Propanol	0.79	160 E J	1.9	380 E
Carbon Disulfide	0.79	0.36 J	2.5	1.1 J
3-Chloropropene	0.79	Not Detected	2.5	Not Detected
Methylene Chloride	0.32	45	1.1	160
Methyl tert-butyl ether	0.16	0.98	0.57	3.5
trans-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Hexane	0.16	0.30	0.56	1.1
1,1-Dichloroethane	0.16	Not Detected	0.64	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.79	0.98	2.3	2.9
cis-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Tetrahydrofuran	0.79	2.6	2.3	7.8
Chloroform	0.16	0.36	0.77	1.8
1,1,1-Trichloroethane	0.16	Not Detected	0.86	Not Detected
Cyclohexane	0.16	0.14 J	0.54	0.47 J
Carbon Tetrachloride	0.16	0.10 J	0.99	0.66 J
2,2,4-Trimethylpentane	0.79	0.14 J	3.7	0.64 J
Benzene	0.16	0.22	0.50	0.70
1,2-Dichloroethane	0.16	Not Detected	0.64	Not Detected
Heptane	0.16	2.1	0.65	8.8
Trichloroethene	0.16	Not Detected	0.85	Not Detected
1,2-Dichloropropane	0.16	Not Detected	0.73	Not Detected
1,4-Dioxane	0.16	Not Detected	0.57	Not Detected
Bromodichloromethane	0.16	Not Detected	1.0	Not Detected
cis-1,3-Dichloropropene	0.16	Not Detected	0.72	Not Detected
4-Methyl-2-pentanone	0.16	50	0.65	210
Toluene	0.16	98 E J	0.60	370 E
trans-1,3-Dichloropropene	0.16	Not Detected	0.72	Not Detected
1,1,2-Trichloroethane	0.16	Not Detected	0.86	Not Detected
Tetrachloroethene	0.16	0.73	1.1	5.0
2-Hexanone	0.16	Not Detected	3.2	Not Detected





Air Toxics

Client Sample ID: B5IA-6 (2014)

Lab ID#: 1412152AR1-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121618r1	Date of Collection:	12/9/14 1:29:00 PM
Dil. Factor:	1.58	Date of Analysis:	12/16/14 08:41 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.16	Not Detected	1.3	Not Detected
1,2-Dibromoethane (EDB)	0.16	Not Detected	1.2	Not Detected
Chlorobenzene	0.16	Not Detected	0.73	Not Detected
Ethyl Benzene	0.16	13	0.69	56
m,p-Xylene	0.16	45	0.69	200
o-Xylene	0.16	4.1	0.69	18
Styrene	0.16	0.54	0.67	2.3
Bromoform	0.16	Not Detected	1.6	Not Detected
Cumene	0.16	Not Detected	0.78	Not Detected
1,1,2,2-Tetrachloroethane	0.16	Not Detected	1.1	Not Detected
Propylbenzene	0.16	Not Detected	0.78	Not Detected
4-Ethyltoluene	0.16	0.11 J	0.78	0.53 J
1,3,5-Trimethylbenzene	0.16	Not Detected	0.78	Not Detected
1,2,4-Trimethylbenzene	0.16	0.32	0.78	1.6
1,3-Dichlorobenzene	0.16	Not Detected UJ	0.95	Not Detected UJ
1,4-Dichlorobenzene	0.16	Not Detected UJ	0.95	Not Detected UJ
alpha-Chlorotoluene	0.79	Not Detected UJ	4.1	Not Detected UJ
1,2-Dichlorobenzene	0.16	Not Detected UJ	0.95	Not Detected UJ
1,2,4-Trichlorobenzene	0.79	Not Detected	5.9	Not Detected
Hexachlorobutadiene	0.79	Not Detected UJ	8.4	Not Detected UJ
Naphthalene	0.79	Not Detected	4.1	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

E = Exceeds instrument calibration range.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	102	70-130
Toluene-d8	108	70-130
4-Bromofluorobenzene	102	70-130





Air Toxics


Client Sample ID: BSIA-7 (2014)

Lab ID#: 1412152AR1-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121619r1	Date of Collection: 12/9/14 2:20:00 PM
Dil. Factor:	1.78	Date of Analysis: 12/16/14 09:39 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.89	0.47 J	4.4	2.3 J
Freon 114	0.18	Not Detected	1.2	Not Detected
Chloromethane	0.89	0.87 J	1.8	1.8 J
Vinyl Chloride	0.18	Not Detected	0.46	Not Detected
1,3-Butadiene	0.18	Not Detected	0.39	Not Detected
Bromomethane	0.89	Not Detected	3.4	Not Detected
Chloroethane	0.89	Not Detected	2.3	Not Detected
Freon 11	0.18	0.20	1.0	1.1
Ethanol	0.89	380 JOE J	1.7	720 JOE
Freon 113	0.18	0.051 J	1.4	0.39 J
1,1-Dichloroethene	0.18	Not Detected	0.70	Not Detected
Acetone	0.89	240 E J	2.1	580 E
2-Propanol	0.89	160 E J	2.2	400 E
Carbon Disulfide	0.89	Not Detected	2.8	Not Detected
3-Chloropropene	0.89	Not Detected	2.8	Not Detected
Methylene Chloride	0.36	65	1.2	230
Methyl tert-butyl ether	0.18	1.5	0.64	5.3
trans-1,2-Dichloroethene	0.18	Not Detected	0.70	Not Detected
Hexane	0.18	0.22	0.63	0.80
1,1-Dichloroethane	0.18	Not Detected	0.72	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.89	0.90	2.6	2.6
cis-1,2-Dichloroethene	0.18	Not Detected	0.70	Not Detected
Tetrahydrofuran	0.89	3.6	2.6	10
Chloroform	0.18	0.34	0.87	1.6
1,1,1-Trichloroethane	0.18	Not Detected	0.97	Not Detected
Cyclohexane	0.18	0.11 J	0.61	0.38 J
Carbon Tetrachloride	0.18	Not Detected	1.1	Not Detected
2,2,4-Trimethylpentane	0.89	0.14 J	4.2	0.64 J
Benzene	0.18	0.23	0.57	0.74
1,2-Dichloroethane	0.18	Not Detected	0.72	Not Detected
Heptane	0.18	2.8	0.73	11
Trichloroethene	0.18	Not Detected	0.96	Not Detected
1,2-Dichloropropane	0.18	Not Detected	0.82	Not Detected
1,4-Dioxane	0.18	Not Detected	0.64	Not Detected
Bromodichloromethane	0.18	Not Detected	1.2	Not Detected
cis-1,3-Dichloropropene	0.18	Not Detected	0.81	Not Detected
4-Methyl-2-pentanone	0.18	71 J	0.73	290
Toluene	0.18	140 E J	0.67	520 E
trans-1,3-Dichloropropene	0.18	Not Detected	0.81	Not Detected
1,1,2-Trichloroethane	0.18	Not Detected	0.97	Not Detected
Tetrachloroethene	0.18	Not Detected	1.2	Not Detected
2-Hexanone	0.89	0.16 J	3.6	0.68 J





Air Toxics

Client Sample ID: B5IA-7 (2014)

Lab ID#: 1412152AR1-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121619r1	Date of Collection: 12/9/14 2:20:00 PM
Dil. Factor:	1.78	Date of Analysis: 12/16/14 09:39 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.18	Not Detected	1.5	Not Detected
1,2-Dibromoethane (EDB)	0.18	Not Detected	1.4	Not Detected
Chlorobenzene	0.18	Not Detected	0.82	Not Detected
Ethyl Benzene	0.18	18	0.77	81
m,p-Xylene	0.18	68	0.77	290
o-Xylene	0.18	6.2	0.77	27
Styrene	0.18	0.85	0.76	3.6
Bromoform	0.18	Not Detected	1.8	Not Detected
Cumene	0.18	Not Detected	0.88	Not Detected
1,1,2,2-Tetrachloroethane	0.18	Not Detected	1.2	Not Detected
Propylbenzene	0.18	Not Detected	0.88	Not Detected
4-Ethyltoluene	0.18	0.13 J	0.88	0.65 J
1,3,5-Trimethylbenzene	0.18	Not Detected	0.88	Not Detected
1,2,4-Trimethylbenzene	0.18	0.41	0.88	2.0
1,3-Dichlorobenzene	0.18	Not Detected UJ	1.1	Not Detected UJ
1,4-Dichlorobenzene	0.18	0.051 J	1.1	0.30 J
alpha-Chlorotoluene	0.89	Not Detected UJ	4.6	Not Detected UJ
1,2-Dichlorobenzene	0.18	Not Detected UJ	1.1	Not Detected UJ
1,2,4-Trichlorobenzene	0.89	Not Detected	6.6	Not Detected
Hexachlorobutadiene	0.89	Not Detected UJ	9.5	Not Detected UJ
Naphthalene	0.89	Not Detected	4.7	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

E = Exceeds instrument calibration range.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	104	70-130
Toluene-d8	109	70-130
4-Bromofluorobenzene	101	70-130





Air Toxics


Client Sample ID: BSIA-2 (2014)

Lab ID#: 1412152AR1-05A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121620r1	Date of Collection:	12/9/14 2:45:00 PM
Dil. Factor:	1.74	Date of Analysis:	12/16/14 10:19 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.87	0.48 J	4.3	2.4 J
Freon 114	0.17	Not Detected	1.2	Not Detected
Chloromethane	0.87	0.89	1.8	1.8
Vinyl Chloride	0.17	Not Detected	0.44	Not Detected
1,3-Butadiene	0.17	0.056 J	0.38	0.12 J
Bromomethane	0.87	Not Detected	3.4	Not Detected
Chloroethane	0.87	Not Detected	2.3	Not Detected
Freon 11	0.17	0.18	0.98	1.0
Ethanol	0.87	320 JOE J	1.6	600 JOE
Freon 113	0.17	0.046 J	1.3	0.35 J
1,1-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Acetone	0.87	210 E J	2.1	510 E
2-Propanol	0.87	150 E J	2.1	380 E
Carbon Disulfide	0.87	4.4	2.7	14
3-Chloropropene	0.87	Not Detected	2.7	Not Detected
Methylene Chloride	0.35	53	1.2	180
Methyl tert-butyl ether	0.17	1.2	0.63	4.2
trans-1,2-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Hexane	0.17	0.20	0.61	0.70
1,1-Dichloroethane	0.17	Not Detected	0.70	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.87	0.96	2.6	2.8
cis-1,2-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Tetrahydrofuran	0.87	2.9	2.6	8.6
Chloroform	0.17	0.39	0.85	1.9
1,1,1-Trichloroethane	0.17	Not Detected	0.95	Not Detected
Cyclohexane	0.17	0.087 J	0.60	0.30 J
Carbon Tetrachloride	0.17	Not Detected	1.1	Not Detected
2,2,4-Trimethylpentane	0.87	0.14 J	4.1	0.67 J
Benzene	0.17	0.39	0.56	1.3
1,2-Dichloroethane	0.17	Not Detected	0.70	Not Detected
Heptane	0.17	2.4	0.71	9.7
Trichloroethene	0.17	Not Detected	0.94	Not Detected
1,2-Dichloropropane	0.17	Not Detected	0.80	Not Detected
1,4-Dioxane	0.17	0.13 J	0.63	0.47 J
Bromodichloromethane	0.17	Not Detected	1.2	Not Detected
cis-1,3-Dichloropropene	0.17	Not Detected	0.79	Not Detected
4-Methyl-2-pentanone	0.17	60	0.71	250
Toluene	0.17	120 E J	0.66	450 E
trans-1,3-Dichloropropene	0.17	Not Detected	0.79	Not Detected
1,1,2-Trichloroethane	0.17	Not Detected	0.95	Not Detected
Tetrachloroethene	0.17	Not Detected	1.2	Not Detected
2-Hexanone	0.17	Not Detected	3.6	Not Detected





Air Toxics

Client Sample ID: B5IA-2 (2014)

Lab ID#: 1412152AR1-05A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121620r1	Date of Collection: 12/9/14 2:45:00 PM
Dil. Factor:	1.74	Date of Analysis: 12/16/14 10:19 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.17	Not Detected	1.5	Not Detected
1,2-Dibromoethane (EDB)	0.17	Not Detected	1.3	Not Detected
Chlorobenzene	0.17	Not Detected	0.80	Not Detected
Ethyl Benzene	0.17	16	0.76	71
m,p-Xylene	0.17	60	0.76	260
o-Xylene	0.17	5.8	0.76	25
Styrene	0.17	0.88	0.74	3.7
Bromoform	0.17	Not Detected	1.8	Not Detected
Cumene	0.17	Not Detected	0.86	Not Detected
1,1,2,2-Tetrachloroethane	0.17	Not Detected	1.2	Not Detected
Propylbenzene	0.17	Not Detected	0.86	Not Detected
4-Ethyltoluene	0.17	0.14 J	0.86	0.70 J
1,3,5-Trimethylbenzene	0.17	Not Detected	0.86	Not Detected
1,2,4-Trimethylbenzene	0.17	0.40	0.86	1.9
1,3-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,4-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
alpha-Chlorotoluene	0.87	Not Detected UJ	4.5	Not Detected UJ
1,2-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,2,4-Trichlorobenzene	0.87	Not Detected	6.4	Not Detected
Hexachlorobutadiene	0.87	Not Detected UJ	9.3	Not Detected UJ
Naphthalene	0.87	Not Detected	4.6	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

E = Exceeds instrument calibration range.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	100	70-130
Toluene-d8	109	70-130
4-Bromofluorobenzene	104	70-130





Air Toxics


Client Sample ID: B5IA-1 (2014)

Lab ID#: 1412152AR1-06A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121621r1	Date of Collection:	12/9/14 2:52:00 PM
Dil. Factor:	3.42	Date of Analysis:	12/16/14 10:54 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	1.7	0.36 J	8.4	1.8 J
Freon 114	0.34	Not Detected	2.4	Not Detected
Chloromethane	1.7	0.80 J	3.5	1.6 J
Vinyl Chloride	0.34	Not Detected	0.87	Not Detected
1,3-Butadiene	0.34	Not Detected	0.76	Not Detected
Bromomethane	1.7	Not Detected	6.6	Not Detected
Chloroethane	1.7	Not Detected	4.5	Not Detected
Freon 11	0.34	0.22 J	1.9	1.2 J
Ethanol	1.7	580 JOE	3.2	1100 JOE
Freon 113	0.34	Not Detected	2.6	Not Detected
1,1-Dichloroethene	0.34	Not Detected	1.4	Not Detected
Acetone	1.7	340 E	4.1	820 E
2-Propanol	1.7	210 E	4.2	510 E
Carbon Disulfide	1.7	Not Detected	5.3	Not Detected
3-Chloropropene	1.7	Not Detected	5.4	Not Detected
Methylene Chloride	0.68	96	2.4	330
Methyl tert-butyl ether	0.34	2.0	1.2	7.4
trans-1,2-Dichloroethene	0.34	Not Detected	1.4	Not Detected
Hexane	0.34	0.21 J	1.2	0.74 J
1,1-Dichloroethane	0.34	Not Detected	1.4	Not Detected
2-Butanone (Methyl Ethyl Ketone)	1.7	0.88 J	5.0	2.6 J
cis-1,2-Dichloroethene	0.34	Not Detected	1.4	Not Detected
Tetrahydrofuran	1.7	5.0	5.0	15
Chloroform	0.34	0.37	1.7	1.8
1,1,1-Trichloroethane	0.34	Not Detected	1.9	Not Detected
Cyclohexane	0.34	Not Detected	1.2	Not Detected
Carbon Tetrachloride	0.34	Not Detected	2.2	Not Detected
2,2,4-Trimethylpentane	1.7	Not Detected	8.0	Not Detected
Benzene	0.34	0.24 J	1.1	0.77 J
1,2-Dichloroethane	0.34	Not Detected	1.4	Not Detected
Heptane	0.34	3.7	1.4	15
Trichloroethene	0.34	Not Detected	1.8	Not Detected
1,2-Dichloropropane	0.34	Not Detected	1.6	Not Detected
1,4-Dioxane	0.34	Not Detected	1.2	Not Detected
Bromodichloromethane	0.34	Not Detected	2.3	Not Detected
cis-1,3-Dichloropropene	0.34	Not Detected	1.6	Not Detected
4-Methyl-2-pentanone	0.34	110	1.4	440
Toluene	0.34	190 E	1.3	730 E
trans-1,3-Dichloropropene	0.34	Not Detected	1.6	Not Detected
1,1,2-Trichloroethane	0.34	Not Detected	1.9	Not Detected
Tetrachloroethene	0.34	Not Detected	2.3	Not Detected
2-Hexanone	0.34	Not Detected	7.0	Not Detected





Air Toxics

Client Sample ID: B5IA-1 (2014)

Lab ID#: 1412152AR1-06A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121621r1	Date of Collection:	12/9/14 2:52:00 PM
Dil. Factor:	3.42	Date of Analysis:	12/16/14 10:54 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.34	Not Detected	2.9	Not Detected
1,2-Dibromoethane (EDB)	0.34	Not Detected	2.6	Not Detected
Chlorobenzene	0.34	Not Detected	1.6	Not Detected
Ethyl Benzene	0.34	26	1.5	110
m,p-Xylene	0.34	96	1.5	420
o-Xylene	0.34	9.1	1.5	40
Styrene	0.34	0.93	1.4	4.0
Bromoform	0.34	Not Detected	3.5	Not Detected
Cumene	0.34	Not Detected	1.7	Not Detected
1,1,2,2-Tetrachloroethane	0.34	Not Detected	2.3	Not Detected
Propylbenzene	0.34	Not Detected	1.7	Not Detected
4-Ethyltoluene	0.34	0.18 J	1.7	0.88 J
1,3,5-Trimethylbenzene	0.34	Not Detected	1.7	Not Detected
1,2,4-Trimethylbenzene	0.34	0.49	1.7	2.4
1,3-Dichlorobenzene	0.34	Not Detected UJ J	2.0	Not Detected UJ
1,4-Dichlorobenzene	0.34	Not Detected UJ J	2.0	Not Detected UJ
alpha-Chlorotoluene	1.7	Not Detected UJ J	8.8	Not Detected UJ
1,2-Dichlorobenzene	0.34	Not Detected UJ J	2.0	Not Detected UJ
1,2,4-Trichlorobenzene	1.7	Not Detected J	13	Not Detected
Hexachlorobutadiene	1.7	Not Detected UJ J	18	Not Detected UJ
Naphthalene	1.7	Not Detected	9.0	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

E = Exceeds instrument calibration range.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	99	70-130
Toluene-d8	108	70-130
4-Bromofluorobenzene	102	70-130





Air Toxics


Client Sample ID: BSIA-AA (2014)

Lab ID#: 1412152AR1-07A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121622r1	Date of Collection:	12/9/14 3:08:00 PM
Dil. Factor:	1.67	Date of Analysis:	12/17/14 05:22 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.84	0.49 J	4.1	2.4 J
Freon 114	0.17	Not Detected	1.2	Not Detected
Chloromethane	0.84	0.85	1.7	1.8
Vinyl Chloride	0.17	Not Detected	0.43	Not Detected
1,3-Butadiene	0.17	Not Detected	0.37	Not Detected
Bromomethane	0.84	Not Detected	3.2	Not Detected
Chloroethane	0.84	Not Detected	2.2	Not Detected
Freon 11	0.17	0.21	0.94	1.2
Ethanol	0.84	3.6 J0 J	1.6	6.7 J0
Freon 113	0.17	0.048 J	1.3	0.37 J
1,1-Dichloroethene	0.17	Not Detected	0.66	Not Detected
Acetone	0.84	5.1	2.0	12
2-Propanol	0.84	2.3	2.0	5.6
Carbon Disulfide	0.84	0.19 J	2.6	0.59 J
3-Chloropropene	0.84	Not Detected	2.6	Not Detected
Methylene Chloride	0.33	0.49	1.2	1.7
Methyl tert-butyl ether	0.17	Not Detected	0.60	Not Detected
trans-1,2-Dichloroethene	0.17	Not Detected	0.66	Not Detected
Hexane	0.17	0.20	0.59	0.70
1,1-Dichloroethane	0.17	Not Detected	0.68	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.84	0.61 J	2.5	1.8 J
cis-1,2-Dichloroethene	0.17	Not Detected	0.66	Not Detected
Tetrahydrofuran	0.84	Not Detected	2.5	Not Detected
Chloroform	0.17	Not Detected	0.82	Not Detected
1,1,1-Trichloroethane	0.17	Not Detected	0.91	Not Detected
Cyclohexane	0.17	0.066 J	0.57	0.22 J
Carbon Tetrachloride	0.17	Not Detected	1.0	Not Detected
2,2,4-Trimethylpentane	0.84	0.10 J	3.9	0.47 J
Benzene	0.17	0.16 J	0.53	0.51 J
1,2-Dichloroethane	0.17	Not Detected	0.68	Not Detected
Heptane	0.17	0.15 J	0.68	0.62 J
Trichloroethene	0.17	Not Detected	0.90	Not Detected
1,2-Dichloropropane	0.17	Not Detected	0.77	Not Detected
1,4-Dioxane	0.17	Not Detected	0.60	Not Detected
Bromodichloromethane	0.17	Not Detected	1.1	Not Detected
cis-1,3-Dichloropropene	0.17	Not Detected	0.76	Not Detected
4-Methyl-2-pentanone	0.17	0.29	0.68	1.2
Toluene	0.17	0.81	0.63	3.0
trans-1,3-Dichloropropene	0.17	Not Detected	0.76	Not Detected
1,1,2-Trichloroethane	0.17	Not Detected	0.91	Not Detected
Tetrachloroethene	0.17	Not Detected	1.1	Not Detected
2-Hexanone	0.17	Not Detected	3.4	Not Detected





Air Toxics

Client Sample ID: B5IA-AA (2014)

Lab ID#: 1412152AR1-07A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121622r1	Date of Collection:	12/9/14 3:08:00 PM
Dil. Factor:	1.67	Date of Analysis:	12/17/14 05:22 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.17	Not Detected	1.4	Not Detected
1,2-Dibromoethane (EDB)	0.17	Not Detected	1.3	Not Detected
Chlorobenzene	0.17	Not Detected	0.77	Not Detected
Ethyl Benzene	0.17	0.20	0.72	0.86
m,p-Xylene	0.17	0.47	0.72	2.0
o-Xylene	0.17	0.14 J	0.72	0.59 J
Styrene	0.17	0.080 J	0.71	0.34 J
Bromoform	0.17	Not Detected	1.7	Not Detected
Cumene	0.17	Not Detected	0.82	Not Detected
1,1,2,2-Tetrachloroethane	0.17	Not Detected	1.1	Not Detected
Propylbenzene	0.17	Not Detected	0.82	Not Detected
4-Ethyltoluene	0.17	0.13 J	0.82	0.66 J
1,3,5-Trimethylbenzene	0.17	0.10 J	0.82	0.51 J
1,2,4-Trimethylbenzene	0.17	0.25	0.82	1.2
1,3-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,4-Dichlorobenzene	0.17	0.030 J	1.0	0.18 J
alpha-Chlorotoluene	0.84	Not Detected UJ	4.3	Not Detected UJ
1,2-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,2,4-Trichlorobenzene	0.84	Not Detected	6.2	Not Detected
Hexachlorobutadiene	0.84	Not Detected UJ	8.9	Not Detected UJ
Naphthalene	0.84	Not Detected	4.4	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	109	70-130
Toluene-d8	106	70-130
4-Bromofluorobenzene	99	70-130





Air Toxics

Client Sample ID: B5IA-4 (2014)

Lab ID#: 1412152AR1-08A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121623r1	Date of Collection: 12/9/14 3:17:00 PM
Dil. Factor:	1.74	Date of Analysis: 12/17/14 06:13 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.87	0.40 J	4.3	2.0 J
Freon 114	0.17	Not Detected	1.2	Not Detected
Chloromethane	0.87	0.92	1.8	1.9
Vinyl Chloride	0.17	Not Detected	0.44	Not Detected
1,3-Butadiene	0.17	Not Detected	0.38	Not Detected
Bromomethane	0.87	Not Detected	3.4	Not Detected
Chloroethane	0.87	Not Detected	2.3	Not Detected
Freon 11	0.17	0.19	0.98	1.1
Ethanol	0.87	360 JOE	1.6	680 JOE
Freon 113	0.17	0.072 J	1.3	0.55 J
1,1-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Acetone	0.87	240 E	2.1	570 E
2-Propanol	0.87	160 E	2.1	400 E
Carbon Disulfide	0.87	Not Detected	2.7	Not Detected
3-Chloropropene	0.87	Not Detected	2.7	Not Detected
Methylene Chloride	0.35	63	1.2	220
Methyl tert-butyl ether	0.17	1.4	0.63	5.0
trans-1,2-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Hexane	0.17	0.42	0.61	1.5
1,1-Dichloroethane	0.17	Not Detected	0.70	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.87	0.84 J	2.6	2.5 J
cis-1,2-Dichloroethene	0.17	Not Detected	0.69	Not Detected
Tetrahydrofuran	0.87	3.5	2.6	10
Chloroform	0.17	0.34	0.85	1.6
1,1,1-Trichloroethane	0.17	Not Detected	0.95	Not Detected
Cyclohexane	0.17	0.094 J	0.60	0.32 J
Carbon Tetrachloride	0.17	Not Detected	1.1	Not Detected
2,2,4-Trimethylpentane	0.87	0.17 J	4.1	0.79 J
Benzene	0.17	0.18	0.56	0.57
1,2-Dichloroethane	0.17	Not Detected	0.70	Not Detected
Heptane	0.17	2.9	0.71	12
Trichloroethene	0.17	Not Detected	0.94	Not Detected
1,2-Dichloropropane	0.17	Not Detected	0.80	Not Detected
1,4-Dioxane	0.17	Not Detected	0.63	Not Detected
Bromodichloromethane	0.17	Not Detected	1.2	Not Detected
cis-1,3-Dichloropropene	0.17	Not Detected	0.79	Not Detected
4-Methyl-2-pentanone	0.17	71 E	0.71	290 E
Toluene	0.17	140 E	0.66	520 E
trans-1,3-Dichloropropene	0.17	Not Detected	0.79	Not Detected
1,1,2-Trichloroethane	0.17	Not Detected	0.95	Not Detected
Tetrachloroethene	0.17	Not Detected	1.2	Not Detected
2-Hexanone	0.17	Not Detected	3.6	Not Detected





Air Toxics

Client Sample ID: B5IA-4 (2014)

Lab ID#: 1412152AR1-08A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	v121623r1	Date of Collection:	12/9/14 3:17:00 PM
Dil. Factor:	1.74	Date of Analysis:	12/17/14 06:13 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.17	Not Detected	1.5	Not Detected
1,2-Dibromoethane (EDB)	0.17	Not Detected	1.3	Not Detected
Chlorobenzene	0.17	Not Detected	0.80	Not Detected
Ethyl Benzene	0.17	20	0.76	85
m,p-Xylene	0.17	Not Detected	0.76	Not Detected
o-Xylene	0.17	6.6	0.76	28
Styrene	0.17	0.80	0.74	3.4
Bromoform	0.17	Not Detected	1.8	Not Detected
Cumene	0.17	Not Detected	0.86	Not Detected
1,1,2,2-Tetrachloroethane	0.17	Not Detected	1.2	Not Detected
Propylbenzene	0.17	0.039 J	0.86	0.19 J
4-Ethyltoluene	0.17	0.18	0.86	0.87
1,3,5-Trimethylbenzene	0.17	Not Detected	0.86	Not Detected
1,2,4-Trimethylbenzene	0.17	0.38	0.86	1.9
1,3-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,4-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
alpha-Chlorotoluene	0.87	Not Detected UJ	4.5	Not Detected UJ
1,2-Dichlorobenzene	0.17	Not Detected UJ	1.0	Not Detected UJ
1,2,4-Trichlorobenzene	0.87	Not Detected	6.4	Not Detected
Hexachlorobutadiene	0.87	Not Detected UJ	9.3	Not Detected UJ
Naphthalene	0.87	Not Detected	4.6	Not Detected

J = Estimated value.

J0 = Estimated value due to bias in the CCV.

E = Exceeds instrument calibration range.

UJ = Analyte associated with low bias in the CCV and/or LCS.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	103	70-130
Toluene-d8	107	70-130
4-Bromofluorobenzene	102	70-130





Air Toxics

Sample Transportation Notice

Relinquishing signature on this document indicates that sample is being shipped in compliance with all applicable local, State, Federal, national, and international laws, regulations and ordinances of any kind. Air Toxics Limited assumes no liability with respect to the collection, handling or shipping of these samples. Relinquishing signature also indicates agreement to hold harmless, defend, and indemnify Air Toxics Limited against any claim, demand, or action, of any kind, related to the collection, handling, or shipping of samples. D.O.T. Hotline (800) 467-4922

180 BLUE RAVINE ROAD, SUITE B
FOLSOM, CA 95630-4719
(916) 985-1000 FAX (916) 985-1020

Page 1 of 1

Project Manager Terry Taylor
Collected by: (Print and Sign) AK
Company AMAI Email _____
Address 110 Corporate Pk City White Plains State NY Zip 10604
Phone 914-391-1225 Fax _____

Project Info:

P.O. # _____

Project # Building S VIProject Name BMS Humacao

Turn Around Time:

☒ Normal☐ Rush

specify _____

Lab Use Only

Pressurized by: _____

Date: _____

Pressurization Gas: _____

N₂ He

Lab I.D.	Field Sample I.D. (Location)	Can #	Date of Collection	Time of Collection	Analyses Requested	Canister Pressure/Vacuum			
						Initial	Final	Receipt	Final (psi)
01A	BSIA-8(2014)	33654	12/9/14	1126	see notes	26.5	6.5		
02A	BSIA-10(2014)	25261	12/9/14	1148	"	30+	9.5		
03A	BSIA-6(2014)	940	12/9/14	1329	"	22.5	5		
04A	BSIA-7(2014)	33889	12/9/14	1420	"	30	8		
05A	BSIA-2(2014)	3737	12/9/14	1445	"	30+	8		
06A	BSIA-1(2014)	12019	12/9/14	1452	"	30	8		
07A	BSIA-AA(2014)	5695	12/9/14	1508	"	29.5	5		
08A	BSIA-4(2014)	25248	12/9/14	1517	"	30+	7"		

Relinquished by: (signature) Date/Time
[Signature] 12/9/14 1520

Received by: (signature) Date/Time
[Signature] 12/10/14 1010

Notes: Acetone, Benzene, Ethylbenzene, toluene, xylene, Methanol, MIBK, Isopropyl Alcohol, Alcohol, Toluene, Methane via ASTM D-1946

Relinquished by: (signature) Date/Time

Received by: (signature) Date/Time

Relinquished by: (signature) Date/Time

Received by: (signature) Date/Time

Lab Use Only	Shipper Name		Air Bill #		Temp (°C)		Condition		Custody Seals Intact?		Work Order #	
	Yes	No	None		Yes	No	None		Yes	No	None	

DATA REVIEW WORKSHEETS

Project Number: 1412152AR1-01A_to_-08A
 Date: 12/09/2014

REVIEW OF VOLATILE ORGANIC PACKAGE

The following guidelines for evaluating volatile organics were created to delineate required validation actions. This document will assist the reviewer in using professional judgment to make more informed decision and in better serving the needs of the data users. The sample results were assessed according to USEPA data validation guidance documents in the following order of precedence: QC criteria from "Compendium Method TO-15. Determination of Volatile Organic Compounds (VOCs) In Air Collected In Specially-Prepared Canisters and Analyzed By Gas Chromatography/Mass Spectrometry (GC/MS), January, 1999"; USEPA Hazardous Waste Support Branch. Validating Air Samples. Volatile Organic Analysis of Ambient Air in Canisters by Method TO-15, (SOP # HW-31. Revision #4. October, 2006). The QC criteria and data validation actions listed on the data review worksheets are from the primary guidance document, unless otherwise noted.

The hardcopied (laboratory name) Eurofins - Air Toxics data package received has been reviewed and the quality control and performance data summarized. The data review for VOCs included:

Lab. Project/SDG No.: 1412152AR1 Sample matrix: Air
 No. of Samples: 8

Trip blank No.: -
 Field blank No.: -
 Equipment blank No.: -
 Field duplicate No.: -

<input checked="" type="checkbox"/> Data Completeness	<input checked="" type="checkbox"/> Laboratory Control Spikes
<input checked="" type="checkbox"/> Holding Times	<input checked="" type="checkbox"/> Field Duplicates
<input checked="" type="checkbox"/> GC/MS Tuning	<input checked="" type="checkbox"/> Calibrations
<input checked="" type="checkbox"/> Internal Standard Performance	<input checked="" type="checkbox"/> Compound Identifications
<input checked="" type="checkbox"/> Blanks	<input checked="" type="checkbox"/> Compound Quantitation
<input checked="" type="checkbox"/> Surrogate Recoveries	<input checked="" type="checkbox"/> Quantitation Limits
<input type="checkbox"/> N/A Matrix Spike/Matrix Spike Duplicate	

Overall Comments: VOCs_by_method_TO-15

Definition of Qualifiers:

J- Estimated results
 U- Compound not detected
 R- Rejected data
 UJ- Estimated nondetect

Reviewer: Rafael Infante
 Date: 07/08/2015

DATA REVIEW WORKSHEETS

DATA COMPLETENESS

MISSING INFORMATION

DATE LAB. CONTACTED

DATE RECEIVED

The form consists of a large rectangular area filled with horizontal lines, intended for data entry. A thick, light-gray diagonal line runs from the top-left corner to the bottom-right corner, crossing out the entire area. This indicates that the data is either missing or has not been reviewed.

DATA REVIEW WORKSHEETS

All criteria were met X
 Criteria were not met
 and/or see below

HOLDING TIMES

The objective of this parameter is to ascertain the validity of the results based on the holding time of the sample from time of collection to the time of analysis.

Complete table for all samples and note the analysis and/or preservation not within criteria

SAMPLE ID	DATE SAMPLED	DATE ANALYZED	pH	ACTION
All samples analyzed within the recommended method holding time				

Criteria

Aqueous samples – 14 days from sample collection for preserved samples (pH ≤ 2, 4°C), no air bubbles.

Aqueous samples – 7 days from sample collection for unpreserved samples, 4°C, no air bubbles.

Soil samples- 7 days from sample collection.

Cooler temperature (Criteria: 4 ± 2 °C): N/A – summa canisters

Actions

If the VOCs vial(s) have air bubbles, estimate positive results (J) and reject nondetects (R).

If the % solids of soil samples is 10-50%, estimates positive results (J) and nondetects (UJ)

If the % solid of soil samples is < 10%, estimate positive results (J) and reject nondetects (R).

If holding times are exceeded but < 14 days beyond criteria, estimate positive results (J) and nondetects (UJ).

If holding times are exceeded but < 28 days beyond criteria, estimate positive results (J) and reject nondetects (R).

If holding times are grossly exceeded (> 28 days beyond criteria), reject all results (R).

If samples were not iced or if the ice were melted (> 10°C), estimate positive results (J) and nondetects (UJ).

DATA REVIEW WORKSHEETS

All criteria were met X
Criteria were not met see below

GC/MS TUNING

The assessment of the tuning results is to determine if the sample instrumentation is within the standard tuning QC limits

 X The BFB performance results were reviewed and found to be within the specified criteria.

 X BFB tuning was performed for every 24 hours of sample analysis.

If no, use professional judgment to determine whether the associated data should be accepted, qualified or rejected.

List the samples affected:

If mass calibration is in error, all associated data are rejected.

DATA REVIEW WORKSHEETS

All criteria were met _____
 Criteria were not met _____
 and/or see below X

CALIBRATION VERIFICATION

Compliance requirements for satisfactory instrument calibration are established to ensure that the instrument is capable of producing and maintaining acceptable quantitative data.

Date of initial calibration: 10/06/14

Dates of continuing calibration: 12/16/14

Instrument ID numbers: MSD-V

Matrix/Level: Air/low

DATE	LAB ID#	FILE	CRITERIA OUT RFs, %RSD, %D, r	COMPOUND	SAMPLES AFFECTED
Initial and continuing calibration met the method performance criteria except for the followings:					
10/06/14	V14L1006D		34 % RSD	1,2,4-trichlorobenzene	All samples.
12/16/14	V121602.d		38 % D	Ethanol	
			38 % D	1,2-dichlorobenzene	
			32 % D	1,3-dichlorobenzene	
			37 % D	1,4-dichlorobenzene	
			33 % D	alpha-chlorotoluene	
			39 % D	Hexachlorobutadiene	

Criteria

All RFs must be > 0.05 regardless of method requirements for SPCC.

All %RSD must be $\leq 15\%$ regardless of method requirements for CCC.

All %Ds must be $\leq 30\%$ regardless of method requirements for CCC.

Method TO-15 does not specify criterion for the curve correlation coefficient (r). A limit for r of ≥ 0.995 has therefore been utilized as professional judgment.

Actions

If any compound has an initial RF or a continuing RF of < 0.05 , estimate positive results (J) and reject nondetects (R), regardless of method requirements.

If any compound has a %RSD $> 15\%$, estimate positive results (J) and use professional judgment to qualify nondetects.

If any compound has a %RSD $> 90\%$, estimate positive results (J) and reject nondetects (R).

If any compound has a % D $> 30\%$, estimate positive results (J) and reject nondetects (R).

If any compound has a % D $> 30\%$, estimate positive results (J) and nondetects (UJ).

If any compound has a % D $> 90\%$, estimate positive results (J) and reject nondetects (R).

If any compound has r < 0.995 , estimate positive results and nondetects.

A separate worksheet should be filled for each initial curve

DATA REVIEW WORKSHEETS

All criteria were met _____
 Criteria were not met _____
 and/or see below X

V A. BLANK ANALYSIS RESULTS (Sections 1 & 2)

The assessment of the blank analysis results is to determine the existence and magnitude of contamination problems. The criteria for evaluation of blanks apply only to blanks associated with the samples, including trip, equipment, and laboratory blanks. If problems with any blanks exist, all data associated with the case must be carefully evaluated to determine whether or not there is an inherent variability in the data for the case, or if the problem is an isolated occurrence not affecting other data.

List the contamination in the blanks below. High and low levels blanks must be treated separately.

Laboratory blanks

DATE ANALYZED	LAB ID	LEVEL/ MATRIX	COMPOUND	CONCENTRATION/ UNITS
---------------	--------	---------------	----------	----------------------

 All method blank meets method specific criteria except for the followings:

<u> 12/16/14 </u>	<u> 1412152AR1-09A </u>	<u> Air/low </u>	<u> Hexane </u>	<u> 0.10 ppbv </u>
			<u> 1,4-dichlorobenzene </u>	<u> 0.038 ppbv </u>
			<u> alpha-chlorotoluene </u>	<u> 0.10 ppbv </u>
			<u> 1,2-dichlorobenzene </u>	<u> 0.046 ppbv </u>
			<u> 1,2,4-trichlorobenzene </u>	<u> 0.12 ppbv </u>
			<u> hexachlorobutadiene </u>	<u> 0.043 ppbv </u>

Note: No action taken, blank concentration < 5 X SQL

 Summa canisters met cleaning certification criteria

Field/Equipment/Trip blank

DATE ANALYZED	LAB ID	LEVEL/ MATRIX	COMPOUND	CONCENTRATION UNITS
---------------	--------	---------------	----------	---------------------

 No field/trip/equipment blanks analyzed with this data package.

DATA REVIEW WORKSHEETS

All criteria were met X
 Criteria were not met
 and/or see below

V B. BLANK ANALYSIS RESULTS (Section 3)

Blank Actions

Action Levels (ALs) should be based upon the highest concentration of contaminant determined in any blank. Do not qualify any blank with another blank. The ALs for samples which have been diluted should be corrected for the sample dilution factor and/or % moisture, where applicable. No positive sample results should be reported unless the concentration of the compound in the samples exceeds the ALs:

ALs = 10x the amount of common contaminants (methylene chloride, acetone, 2-butanone, and toluene)

ALs = 5x for any other compounds

Specific actions are as follows:

If the concentration is < sample quantitation limit (SQL) and \leq AL, report the compound as not detected (U) at the SQL.

If the concentration is \geq SQL but \leq AL, report the compound as not detected (U) at the reported concentration.

If the concentration is \geq SQL and $>$ AL, report the concentration unqualified.

Notes:

High and low level blanks must be treated separately

Compounds qualified "U" for blank contamination are still considered "hits" when qualifying for calibration criteria.

CONTAMINATION SOURCE/LEVEL	COMPOUND	CONC/UNITS	AL/UNITS	SQL	AFFECTED SAMPLES

DATA REVIEW WORKSHEETS

All criteria were met X
 Criteria were not met
 and/or see below

SURROGATE SPIKE RECOVERIES

Laboratory performance of individual samples is established by evaluation of surrogate spike recoveries. All samples are spiked with surrogate compounds prior to sample analysis. The accuracy of the analysis is measured by the surrogate percent recovery. Since the effects of the sample matrix are frequently outside the control of the laboratory and may present relatively unique problems, the validation of data is frequently subjective and demands analytical experience and professional judgment.

List the percent recoveries (%Rs) which do not meet the criteria for surrogate recovery.

Matrix: solid/aqueous

SAMPLE ID	SURROGATE COMPOUND			ACTION
	1,2-DICHLOROETHANE- d4	Toluene- d8	4-BFB	

 Surrogate recoveries within laboratory control limits

QC Limits* (Air)

 LL to UL 70 to 130 70 to 130 70 to 130

- * QC limits are laboratory in-house performance criteria, LL = lower limit, UL = upper limit.
- * If QC limits are not available, use limits of 80 – 120 % for aqueous and 70 – 130 % for solid samples.

Actions:

QUALITY	%R < 10%	%R = 10% - LL	%R > UL
Positive results	J	J	J
Nondetects results	R	UJ	Accept

Surrogate action should be applied:

If one or more surrogate in the VOC fraction is out of specification, but has a recovery of > 10%.

If any one surrogate in a fraction shows < 10 % recovery.

DATA REVIEW WORKSHEETS

All criteria were met _____
 Criteria were not met _____
 and/or see below ___N/A___

VII. A MATRIX SPIKE/MATRIX SPIKE DUPLICATE (MS/MSD)

This data is generated to determine long term precision and accuracy in the analytical method for various matrices. This data alone cannot be used to evaluate the precision and accuracy of individual samples. If any % R in the MS or MSD falls outside the designated range, the reviewer should determine if there are matrix effects, i.e. LCS data are within the QC limits but MS/MSD data are outside QC limit.

1. MS/MSD Recoveries and Precision Criteria

The laboratory should use one MS and a duplicate analysis of an unspiked field sample if target analytes are expected in the sample. If target analytes are not expected, MS/MSD should be analyzed.

List the %Rs, RPD of the compounds which do not meet the criteria.

Sample ID: _____ Matrix/Level: _____

MS OR MSD	COMPOUND	% R	RPD	QC LIMITS	ACTION
MS/MSD are not required as part of Method TO-15; blank spike used to assess accuracy					

- * QC limits are laboratory in-house performance criteria, LL = lower limit, UL = upper limit.
- * If QC limits are not available, use limits of 70 – 130 %.

Actions:

QUALITY	%R < LL	%R > UL
Positive results	J	J
Nondetects results	R	Accept

MS/MSD criteria apply only to the unspiked sample, its dilutions, and the associated MS/MSD samples:

If the % R for the affected compounds were < LL (or 70 %), qualify positive results (J) and nondetects (UJ).

If the % R for the affected compounds were > UL (or 130 %), only qualify positive results (J).

If 25 % or more of all MS/MSD %R were < LL (or 70 %) or if two or more MS/MSD %Rs were < 10%, qualify all positive results (J) and reject nondetects (R).

A separate worksheet should be used for each MS/MSD pair.

DATA REVIEW WORKSHEETS

All criteria were met _____
Criteria were not met _____
and/or see below N/A

VII. B MATRIX SPIKE/MATRIX SPIKE DUPLICATE

MS/MSD – Unspiked Compounds

It should be noted that Method TO-15 does not specify a MS/MSD criteria for the unspiked compounds in the sample. A %RSD of < 50% has therefore been utilized as professional judgment.

If all target analytes were spiked in the MS/MSD, this review element is not applicable.

List the %RSD of the compounds which do not meet the criteria.

Sample ID: _____ Matrix/Level/Unit: _____

[illegible]

Actions:

- * If the % RSD > 50, qualify the positive result in the unspiked samples as estimated (J).
* If the % RSD is not calculated (NC) due to nondetected value, use professional judgment to qualify the data.

All criteria were met X
 Criteria were not met
 and/or see below

VIII. LABORATORY CONTROL SAMPLE (LCS) ANALYSIS

This data is generated to determine accuracy of the analytical method for various matrices.

1. LCS Recoveries Criteria

Where LCS spiked with the same analyte at the same concentrations as the MS/MSD?
 Yes or No. If no make note in data review memo.

List the %R of compounds which do not meet the criteria

LCS ID	COMPOUND	% R	QC LIMIT
LCS/LCSD_% recoveries_and_RPD_within_laboratory_control_limits_except_for_the followings:			
1412152AR1-11A/-11AA	1,2,4-trimethylbenzene	143_%/142_%	70_-130
	alpha-chlorotoluene	167_%/159_%	70_-130
	Ethanol	149_%/152_%	70_-130

Note: Sample results qualified accordingly in affected sample.

- * QC limits are laboratory in-house performance criteria, LL = lower limit, UL = upper limit.
- * If QC limits are not available, use limits of 70 – 130 %.

Actions:

QUALITY	%R < LL	%R > UL
Positive results	J	J
Nondetects results	R	Accept

All analytes in the associated sample results are qualified for the following criteria.

If 25 % of the LCS recoveries were < LL (or 70 %), qualify all positive results (j) and reject nondetects (R).

If two or more LCS were below 10 %, qualify all positive results as (J) and reject nondetects (R).

2. Frequency Criteria:

Where LCS analyzed at the required frequency and for each matrix? Yes or No.

If no, the data may be affected. Use professional judgment to determine the severity of the effect and qualify data accordingly. Discuss any actions below and list the samples affected.

All criteria were met X
 Criteria were not met
 and/or see below _____

IX. LABORATORY DUPLICATE PRECISION

Sample IDs: LCS/LCSD Matrix: Air

Laboratory duplicates samples may be taken and analyzed as an indication of overall precision. These analyses measure both field and lab precision; therefore, the results may have more variability than laboratory duplicates which only laboratory performance. It is also expected that soil duplicate results will have a greater variance than water matrices due to difficulties associated with collecting identical field duplicate samples.

The project QAPP should be reviewed for project-specific information.

Suggested criteria: RPD \pm 25% for air samples. If both samples and duplicate are <5 SQL, the RPD criteria is doubled.

COMPOUND	SQL	SAMPLE CONC.	DUPLICATE CONC.	RPD	ACTION
RPD within laboratory and generally acceptable control limits.					

Actions:

Qualify as estimated positive results (J) and nondetects (UJ) for the compound that exceeded the above criteria. For organics, only the sample and duplicate will be qualified.

If an RPD cannot be calculated because one or both of the sample results is not detected, the following actions apply:

If one sample result is not detected and the other is greater than 5x the SQL qualify (J/UJ).

If one sample value is not detected and the other is greater than 5x the SQL and the SQLs for the sample and duplicate are significantly different, use professional judgment to determine if qualification is appropriate.

If one sample value is not detected and the other is less than 5x, use professional judgment to determine if qualification is appropriate.

If both sample and duplicate results are not detected, no action is needed.

DATA REVIEW WORKSHEETS

All criteria were met ___ N/A ___
Criteria were not met
and/or see below _____

IX. FIELD DUPLICATE PRECISION

Sample IDs: _____ - _____

Matrix: _____

Field duplicate samples may be taken and analyzed as an indication of overall precision. These analyses measure both field and lab precision; therefore, the results may have more variability than laboratory duplicates which only laboratory performance. It is also expected that soil duplicate results will have a greater variance than water matrices due to difficulties associated with collecting identical field duplicate samples.

The project QAPP should be reviewed for project-specific information.

Suggested criteria: RPD \pm 25% for air samples. If both samples and duplicate are <5 SQL, the RPD criteria is doubled.

COMPOUND	SQL	SAMPLE CONC.	DUPLICATE CONC.	RPD	ACTION
No field duplicate analyzed with this data package, LCS/LCSD results used to assess precision. RPD within generally acceptable control limits.					

Actions:

Qualify as estimated positive results (J) and nondetects (UJ) for the compound that exceeded the above criteria. For organics, only the sample and duplicate will be qualified.

If an RPD cannot be calculated because one or both of the sample results is not detected, the following actions apply:

If one sample result is not detected and the other is greater than 5x the SQL qualify (J/UJ).

If one sample value is not detected and the other is greater than 5x the SQL and the SQLs for the sample and duplicate are significantly different, use professional judgment to determine if qualification is appropriate.

If one sample value is not detected and the other is less than 5x, use professional judgment to determine if qualification is appropriate.

If both sample and duplicate results are not detected, no action is needed.

DATA REVIEW WORKSHEETS

All criteria were met X
 Criteria were not met
 and/or see below

X. INTERNAL STANDARD PERFORMANCE

The assessment of the internal standard (IS) parameter is used to assist the data reviewer in determining the condition of the analytical instrumentation.

List the internal standard area of samples which do not meet the criteria.

- * Area of +40% or -40% of the IS area in the associated calibration standard.
- * Retention time (RT) within ± 0.06 seconds of the IS area in the associated calibration standard.

DATE	SAMPLE ID	IS OUT	IS AREA	ACCEPTABLE RANGE	ACTION
------	-----------	--------	---------	------------------	--------

Internal standard area and retention times within laboratory control limits for both samples and calibration standards

Actions:

1. IS actions should be applied to the compound quantitated with the out-of-control ISs

QUALITY	IS AREA < -40%		IS AREA > + 40%
Positive results	J		J
Nondetected results	R		ACCEPT

2. If a IS retention time varies more than 0.330 seconds, the chromatographic profile for that sample must be examined to determine if any false positive or negative exists. For shifts of a large magnitude, the reviewer may consider partial or total rejection of the data for the sample fraction.

DATA REVIEW WORKSHEETS

All criteria were met X
Criteria were not met
and/or see below

XII. SAMPLE QUANTITATION

The sample quantitation evaluation is to verify laboratory quantitation results. In the space below, please show a minimum of one sample calculation:

1412152AR1-01A

Ethanol RF = 1.02216

$$[] = (728110)(5.0)/(89897)(1.02216)$$

$$= 39.62 \text{ ppbv OK}$$

DATA REVIEW WORKSHEETS

All criteria were met X
 Criteria were not met
 and/or see below

XII. QUANTITATION LIMITS

A. Dilution performed

SAMPLE ID	DILUTION FACTOR	REASONS FOR DILUTION
Dilution was performed on samples by a factor of less than 2 except the following:		
1412152AR1-06A	3.42	High levels of target species.

B. Percent Solids

List samples which have $\leq 50\%$ solids

Actions:

If the % solids of a soil sample is 10-50%, estimate positive results (J) and nondetects (UJ)

If the % solids of a soil sample is $< 10\%$, estimate positive results (J) and reject nondetects (R)